


Future perspectives in quality assurance

Piet Meijer
ECAT Foundation
The Netherlands



Regular EQA approach Consensus value



	N	Assigned Value (IU/dL)	CV (%)	Range
Total group	280	23	11.5	13 - 36



Lab result : 26 IU/dL

Z- score : $(26-23)/2.65 = 1.13$



% Deviation : $[(26-23)/23]*100\% = 13.0\%$




Is this approach suitable for performance assessment with extended half-life product samples?



Example: Alprolix (60 IU/dL)





Target : 60 IU/dL
 Consensus : 49.8 IU/dL
 Result : 60 IU/dL
 Result : 40 IU/dL

% Deviation	40 IU/dL	60 IU/dL
Consensus	- 19.7%	20.5%
Target	- 33.3%	0.0%



EQA approach for drug monitoring

- Performance assessment with respect to the target value
- Drug recovery
- Target value on the basis of potency labelling



Questionnaire on interest for EQA surveys for extended half-life FVIII and FIX products.

Number of responders : 157

- 27% is interested in specific surveys for EHL products.
- 65% prefer product-specific surveys / 35% unknown samples





Validation

Confirmation, through the provision of objective evidence, that the requirements for a specific intended use or application have been fulfilled

ISO 15189







Assay Project BAY 94-9027



QC Sample	Target (IU/mL)	Actual Value (IU/mL)	Uncertainty	
			Low	High
1	0.05	0.05	0.048	0.054
2	0.1	0.12	0.113	0.123
3	0.25	0.26	0.242	0.282
4	0.5	0.49	0.454	0.526
5	1.5	1.40	1.290	1.506



Assay Project BAY94-9027

You are here: [Home](#) > [Information](#) > [Assay Project BAY94-9027](#)

· [POCT Single kits](#)

· [POCT Single kits – Order form](#)

· [EQA General Information](#)

· [EQA Programme Manual](#)

· [EQA Programme and prices 2018](#)

· [EQA Programme and prices 2019](#)

· [EQA Survey Schedule 2018](#)

· [EQA Survey Schedule 2019](#)

· [EQA Registration](#)

· [Assay Project BAY94-9027](#)

· [Accreditation](#)

· [Terms of Delivery](#)

Assay Project BAY94-9027:



BAY 94-9027 is a Site-Specifically PEGylated B-Domain-Deleted Recombinant FVIII which has demonstrated safety and efficacy in more than 5 years of clinical studies.

Because BAY 94-9027 is a modified molecule, there is the potential for differences in performance when it is measured in certain coagulation assays established for standard FVIII products. To support medical laboratories in their need for BAY 94-9027 assay performance information, Bayer has entered into a partnership with North American Specialized Coagulation Laboratory Association (NASCOLA) and External quality Control for Assays and Tests (ECAT) to develop BAY 94-9027 quality control samples for proficiency evaluation and further data collection on assay performance in the coagulation testing community.

The Evaluation Sample Sets contain BAY 94-9027 active pharmaceutical ingredient (API) spiked into FVIII depleted human plasma at 5 different target concentrations from 0.05 IU/mL to 1.5 IU/mL. The individual BAY 94-9027 QC sample vials are lyophilized for reconstitution with 2 mL of water prior to potency testing. Laboratories are requested to measure these samples in multiple dilutions with either ne-stage clotting assay and/or a chromogenic assay according to their regular laboratory protocol. Results and detailed information about the methodology used are returned to ECAT using a standard report form. Laboratories will receive a report after evaluation of their results.

Laboratories interested in participation in this project can obtain one Evaluation Set free-of-charge.

Instructions for testing and reporting of data are included in shipment.

To order a BAY 94-9027 Evaluation set complete to order form:

Order form BAY94-9027 Evaluation Set

This is the order form for the BAY94-9027 Evaluation Set.

* mandatory field



Future perspectives in quality assurance:

- Surveys for drug monitoring
- Support in method validation
- Education

